COMMENTARY

Open Access

Designing gene panels for tumor mutational burden estimation: the need to shift from 'correlation' to 'accuracy'

Hao-Xiang Wu^{1†}, Zi-Xian Wang^{1†}, Qi Zhao², Feng Wang^{1*} and Rui-Hua Xu^{1*}

Abstract

Tumor mutational burden (TMB) assessment is at the forefront in precision medicine. The TMB could represent a biomarker for immune checkpoint inhibitors (ICIs) responses. Whole exome sequencing (WES) is the gold standard to derive the TMB; while targeted next-generation sequencing panels might be more feasible. However, mainstream panels use 'correlation' (R^2) between panel- and WES-based TMB to validate TMB estimation, which could be vulnerable to be distorted by cases with relatively ultra-high TMB within each cancer type. The FDA-approved FoundationOne CDx (F1CDx) panel-based TMB estimation seemed reliable ($R^2 \geq 0.75$) in 24 out of 33 cancer types from the Cancer Genome Atlas, but most of them were overestimated by correlation as only seven cancer types had satisfactory accuracy (the proportion of cases correctly identified as TMB-high or TMB-low using panel-based TMB) above 90%. After removing cases with relatively ultra-high TMB within each cancer type, the correlation (R^2) in 16 of these 24 cancer types declined dramatically ($\Delta > 0.25$) while all of their accuracy remained generally constant, indicating that accuracy is more robust than correlation. Similar results were also observed in other four panels. Further incorporating accuracy in panel design revealed that the minimal number of genes needed to achieve $\geq 90\%$ accuracy varied among cancer types and correlated negatively with their TMB levels ($p = 0.001$). In summary, currently available panels can accurately assess TMB only in several particular cancer types; and accuracy outperformed correlation in assessing the performance of panel-based TMB estimation. Accuracy and cancer type individualization should be incorporated in designing panels for TMB estimation.

Keywords: Accuracy, Correlation, Panel, TCGA, Tumor mutational burden

Background

The tumor mutational burden (TMB) is increasingly recognized as a potential biomarker for the response to immune checkpoint inhibitors (ICIs). Clinical studies have noted the association of high TMB with improved patient responses and survival benefit after ICI treatment either in a single cancer type (eg. non-small-cell lung cancer [1], melanoma [2], gastric cancer [3], and urothelial cancer [4]) or in combined cohort of multiple cancer types [5]. And the application of TMB as a

biomarker for ICI treatment is now being prospectively tested [6, 7]. Therefore, TMB assessment has become a research hot spot in the field of precision medicine.

Currently, whole exome sequencing (WES)-derived TMB values are considered as the gold standard, but the high cost and long turnaround time limit the routine diagnostic applicability of WES. Therefore, targeted next-generation sequencing (NGS) panels have been promoted as a simpler and cheaper approach for TMB estimation [8]. Both the FDA-approved FoundationOne CDx (F1CDx) panel and the FDA-authorized MSK-IMPACT panel used 'correlation' (R^2) between panel- and WES-based TMB to validate the capability of panel-based TMB estimation, and it's claimed that these panels can assess TMB accurately ($R^2 = 0.74$ for F1CDx and $R^2 = 0.76$ for MSK-IMPACT) [9, 10]. Furthermore, Wang and colleagues recently claimed that a panel with more than 150 genes was

* Correspondence: wangfeng@sysucc.org.cn; xurh@sysucc.org.cn

[†]Hao-Xiang Wu and Zi-Xian Wang contributed equally to this work.

[‡]Feng Wang and Rui-Hua Xu are joint senior authors.

¹Department of Medical Oncology, State Key Laboratory of Oncology in South China, Collaborative Innovation Center for Cancer Medicine, Sun Yat-Sen University Cancer Center, Guangzhou 510060, People's Republic of China

Full list of author information is available at the end of the article

© The Author(s). 2019 **Open Access** This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated.

sufficient for accurate TMB estimation based on their findings that the ‘correlation’ (R^2) between the panel- and WES-based TMB gradually increased along with a decreasing standard deviation and reached a plateau after 150 genes were included [11]. However, the overall correlation between the panel- and WES-based TMB could be substantially distorted by outliers (i.e. cases with relatively ultra-high TMB within each cancer type) [12], which might lead to overestimation of the reliability of TMB estimation. Therefore, the present study aimed to assess the reliability of TMB estimation using these panels across multiple cancer types; and compared the robustness of ‘correlation’ and ‘accuracy’ in assessing the performance of panel-based TMB estimation.

Methods

Ten thousand one hundred forty-seven cases across 33 cancer types from the Cancer Genome Atlas (TCGA) were included in this study. For WES mutation data, we used the uniform somatic called variants determined by TCGA MC3 project, which were comprehensively curated from detection using seven methods (MuTect, MuSE, VarScan2, Radia, Pindel, Somatic Sniper, Indelocator) [13]. The TMB was calculated as the number of non-synonymous somatic, coding, base substitution, and indel mutations per megabase (Mb) of genome examined, and 38 Mb was used as the estimate of the whole exome size [10]. Five currently available NGS panels for TMB determination (i.e. F1CDx, MSK-IMPACT, Illumina TSO500, Oncomine TML, QIAseq TMB) were investigated, and in silico simulated panel-based TMB scores were calculated by dividing the number of somatic mutations in the targeted genes by the region captured of corresponding panels per manufacturers’ instructions. Notably, for F1CDx and TSO500, synonymous mutations were also included in order to reduce sampling noise as the developers proposed. Previous studies have suggested that the inclusion of synonymous mutations could enhance the precision of panel-based TMB estimation [11, 14]. Still we preferred to retain the original algorithm for the other three panels without inclusion of synonymous mutations. We should also notice that although these panels have been developed for TMB determination, all of them, except for F1CDx and MSK-IMPACT, have not been approved by FDA as diagnostic assay and are still for research use only.

The primary outcomes were ‘correlation’ and ‘accuracy’. Correlations between panel- and WES-based TMB were examined using the Pearson correlation coefficient (R^2). Accuracy was calculated as the proportion of cases that were correctly identified as either high TMB or low TMB using panel-based TMB. Besides accuracy, we additionally calculated the false positive rate (proportion of cases misclassified as TMB-high), false negative rate (proportion of cases misclassified as TMB-low), positive

percentage agreement (calculated by dividing the number of true TMB-high by the total sum of true TMB-high and false TMB-low) and negative percentage agreement (calculated by dividing the number of true TMB-low by the sum of all true TMB-low and false TMB-high). Although retrospective analyses have established the predictive function of high TMB for a better response to ICIs, the optimal cutpoint to define high TMB varied among studies [15]. Based on the results of a multi-cancer cohort receiving ICI treatment, Samstein and colleagues proposed that there may not be one universal definition of high TMB; while the top 20% in each cancer type may serve as an option [5]. Thus, the top 20% in each cancer type was used as the cutpoint to define high TMB, and we varied the cutpoint from the top 10–50% for additional analysis.

As ‘correlation’ would be substantially distorted by cases with relatively ultra-high TMB (defined as cases with TMB ranking top 5% within a particular cancer type), we test the robustness of correlation and accuracy by successively removing cases with WES-based TMB ranking from the top 1–5% in each cancer type. In addition, we also examined the correlation between panel- and WES-based TMB in different TMB subgroups (top 5%, top 5–20%, and bottom 80%).

To explore the minimal number of genes needed for accurate TMB estimation (accuracy $\geq 90\%$) in each cancer type, we randomly extracted genes within the genomic scope to constitute randomized panels with size ranging from 150 to 1000 genes. The genes included in each size of panel were extracted randomly 1000 times. The minimal number of genes was truncated at 1000 for cancer types that needed more than 1000 genes to obtain a mean accuracy $\geq 90\%$.

Results and discussion

The correlations between the F1CDx- and WES-based TMB across 33 cancer types are shown in Fig. 1a, top panel. In accordance with previous studies [10], F1CDx seemed to accurately assess TMB ($R^2 \geq 0.75$) in at least 24 out of 33 cancer types. However, when using the top 20% in each cancer type as the cutpoint to define high TMB, the accuracy of these 24 cancer types ranged largely from 56 to 99% (Fig. 1a, bottom panel), and only seven cancer types had satisfactory accuracy ($\geq 90\%$), while the false positive and false negative rate were considerable in other cancer types (Fig. 1a, bottom panel). Besides, the positive percentage agreement was below 80% in more than two-thirds of the 33 cancer types while the negative percentage agreements were rather high compared with corresponding positive percentage agreements (Fig. 1b). These results indicated that F1CDx-based TMB estimation was only reliable in particular cancer types (e.g. cervical squamous-cell carcinoma and

endocervical adenocarcinoma [CESC], colon adenocarcinoma [COAD], head and neck squamous cell carcinoma [HNSC], lung adenocarcinoma [LUAD], skin cutaneous melanoma [SKCM], stomach adenocarcinoma [STAD], and uterine corpus endometrial carcinoma [UCEC]); while the reliability of F1CDx-based TMB estimation was overestimated by correlation in the other 17 cancer types with $R^2 \geq 0.75$. If we classified patients into TMB-high and TMB-low subgroups according to F1CDx-based TMB estimation in these 17 cancer types, considerable misclassification would happen, and mainly due to the misclassification of TMB-low patients as TMB-high (false positive).

The reason why the reliability of F1CDx-based TMB estimation was overestimated by correlation is that correlation is vulnerable to be distorted by the common presence of cases with relatively ultra-high TMB within each cancer type (Additional file 1: Figure S1). For example, there were 177 pancreatic adenocarcinoma (PAAD) cases in total, whose distribution of TMB (median [IQR], 0.92 [0.60–1.23] Mut/Mb) was presented in Additional file 1: Figure S1. Among these 177 PAAD cases, the F1CDx- and WES-based TMB estimation were highly correlated ($R^2 = 1.00$). But if a relatively ultra-hypermutated case (TCGA-IB-7651) was omitted, the panel-based TMB estimation within the remaining 176 PAAD cases was found to be quite inaccurate and the correlation (R^2) declined sharply to 0.33 (Additional file 1: Figure S2).

Therefore, we further tested the robustness of ‘correlation’ in assessing the performance of panel-based TMB estimation by removal of cases with relatively ultra-high TMB (defined as cases with TMB ranking top 5% within a particular cancer type). After successively removing the cases with WES-based TMB ranking from the top 1–5% in each of the 24 cancer types with $R^2 \geq 0.75$, a dramatic decline in correlation ($\Delta > 0.25$) between F1CDx- and WES-based TMB was observed in more than half (16/24) of them (Fig. 2a).

In contrast, we successively removed the cases with WES-based TMB ranking from the top 1–5% in each cancer type, and found that the accuracy was generally constant in all 24 cancer types compared with correlation (Fig. 2a). Similar results were observed when we varied the cutpoint from the top 10–50% in each cancer type to define high TMB (Additional file 1: Figure S3). The superiority of accuracy over correlation was also prominent in other currently available NGS panels, e.g. MSK-IMPACT, Illumina TSO500, OncoPrint TML, and QIAseq TMB panel (Additional file 1: Figure 4A–D). One may concern that the removal of top 5% cases systematically removed cases only from the TMB-high group and reduced the sample size of this group to 75%. Therefore, we also retained the top 5% cases and examined the correlation between F1CDx- and WES-based

TMB in different TMB subgroups (top 5%, top 5–20%, and bottom 80%). As shown in Fig. 2b, in more than two-thirds of 24 cancer types with $R^2 \geq 0.75$, the correlations between F1CDx- and WES-based TMB estimation in the top 5–20% subgroup and the bottom 80% subgroup were similar, but both were lower than that in the top 5% subgroup and the total cases, indicating that the correlation was distorted by cases with relatively ultra-high TMB and the reliability of panel-based TMB estimation was actually suboptimal in these cancer types. Additional analysis using the other four NGS panels confirmed this result (Additional file 1: Figure 5A–D).

These results strongly suggested that accuracy was a robust and better index compared with correlation in assessing the performance of panel-based TMB estimation and could be readily incorporated into the design of panels for TMB estimation.

In previous studies, the performance of panel-based TMB estimation were examined within limited sample size ($n = 29$ for F1CDx and $n = 106$ for MSK-IMPACT) [9, 10]. Based on the *in silico* analysis of well-curated WES data from more than 10,000 cases, we found that the precision of F1CDx or other panels-based TMB estimation might not be adequate in all cancer types, especially in those with intermediate to low TMB levels (Additional file 1: Figure S6), in most of which the reliability of panel-based TMB was otherwise overestimated by ‘correlation’. Therefore we should not validate panels for TMB estimation simply relying on ‘correlation’ as sometimes it could give rise to misleading results, which would probably cause improper application of ICIs. Besides, the accuracy of F1CDx-based TMB estimation varied among cancer types and positively correlated with their TMB levels (Additional file 1: Figure S6, $p < 0.001$), indicating that one universal NGS panel may not be enough for TMB estimation across multiple cancer types, while cancer type individualized panels accounting for their TMB levels could be more applicable.

‘Correlation’ is a measure of the linear relationship between two variables and can be readily interpreted. For assessment of panels, the higher the correlation is, the more precise the panel would be. But it is often distorted by cases with relatively ultra-high TMB, which cannot be avoided in most cancer types. While ‘accuracy’ is based on the method of dichotomy, thus it would not be significantly affected by outliers, and is more applicable in clinical settings. Besides the precise estimation of TMB values, we would focus more on how many patients will be incorrectly classified as TMB-high (false positive) and TMB-low (false negative). An accuracy of 90% (e.g. HNSC) means that 10% of the patients would be misclassified according to the results of F1CDx and consequently leads to improper decisions on the application of ICIs in these patients. Meanwhile, as a method of dichotomy, ‘accuracy’

Fig. 2 Accuracy outperformed correlation in assessing the performance of panel-based tumor mutational burden (TMB) estimation. **a** Changes in the correlation and accuracy when successively removing the cases with WES-based TMB ranking from the top 1–5% in each cancer type using the F1CDx panel. **b** The correlation between F1CDx- and WES-based TMB in different TMB subgroups (top 5%, top 5–20%, and bottom 80%) in 24 cancer types with $R^2 \geq 0.75$. **c** The minimal number of genes needed to obtain a mean accuracy $\geq 90\%$ varied among cancer types and correlated negatively with their TMB levels

focuses more on the proportion of misclassification rather than the exact TMB estimation of every sample; and the cutpoint needs to be prespecified. Although the continuum of TMB values also matters as the survival benefit was more pronounced when TMB cutpoint got higher [5], definite cutpoints for TMB may be more practical and interpretable in clinical settings. Another effective biomarker for ICI treatment, PD-L1 expression, also obtained approval based on definite cutpoints. Thus the better option might be carefully combining ‘accuracy’ along with ‘correlation’ in the validation of NGS panels.

To give an example of incorporating accuracy into the design of panels for TMB estimation, we randomly extracted genes to generate virtual panels of 150 to 1000 genes to explore the minimal number of genes needed for accurate TMB estimation. In most cancer types, the mean accuracy gradually increased but few reached a plateau (Additional file 1: Figure S7). The minimal number of genes needed to obtain a mean accuracy above 90% varied among cancer types (median [range], 940 [150–1000]) and correlated negatively with their TMB levels (Fig. 2c, $p = 0.001$), which was in line with our proposal that panels for TMB estimation should be cancer type individualized in terms of cost and benefit. For cancer types with higher TMB levels, smaller panels are sufficient to capture the mutational burden, while for cancer types with lower TMB levels, larger panels are needed. Certainly, randomly selected gene panels may not be appropriate for TMB estimation, and it's not cost-effective to develop a NGS panel only for TMB estimation. As mutational spectrum is divergent across cancer types, cancer type individualized panels in which the size (how many genes) and composition (what genes) are carefully elaborated would be more applicable both for TMB estimation and identification of actionable targets.

It's thought-provoking that the first prospective clinical trial (CheckMate 227) seems to fail in establishing the predictive function of TMB, probably due to the dilution of treatment effect caused by misclassification of TMB-high and TMB-low patients by F1CDx-based TMB estimation. The key usage of ‘accuracy’ is that it is robust in assessing the reliability of panel-based TMB albeit the common presence of outliers. As using NGS panels to determine TMB is more feasible than WES, a panel with high accuracy could reduce the misclassification in clinical trials, thus guarantees greater power in detecting the predictive function of TMB and establishes validated TMB cutpoints.

A major limitation of this study is that there are still lots of pre-analytic issues about the clinical application of panel-based TMB. For instance, the variation of sample storage time, the high scoring failure rate, and so on [16]. Therefore, incorporating the methods of ‘accuracy’

and ‘cancer type individualization’ in panel design requires wet-lab validation before it could be used in clinical practice.

Conclusions

Increasing numbers of clinical trials include the TMB as a key design component; therefore, accurate TMB assessment is fundamental to ensure reliable and reproducible identification of those patients likely to benefit from ICI treatment. The present study showed that, the currently available NGS panels can assess TMB accurately only in several particular cancer types; and with the presence of cases with relatively ultra-high TMB, ‘correlation’ is unreliable to evaluate the performance of panel-based TMB estimation in most cancer types, whereas ‘accuracy’ is a superior index in this situation. Furthermore, cancer type individualized panels might be a better strategy to guarantee robust TMB estimation and thus greater power in prospectively detecting the predictive function of TMB across multiple cancer types.

Additional file

Additional file 1: Figure S1. The tumor mutational burden across 33 cancer types in TCGA. **Figure S2.** The correlation between F1CDx- and WES-based TMB in PAAD after the removal of a relatively ultra-hypermutated case. **Figure S3.** Changes in the accuracy with cutpoint varying from the top 10–50% when successively removing the cases with WES-based TMB ranking from the top 1–5% in each cancer type. **Figure S4.** Changes in the correlation and accuracy when successively removing the cases with WES-based TMB ranking from the top 1–5% in each cancer type using the MSK-IMPACT panel (A), the Illumina TSO500 panel (B), the OncoPrint TML panel (C), and the QIAseq TMB panel (D). **Figure S5.** The correlation between panel- and WES-based TMB in different TMB subgroups (top 5%, top 5–20%, bottom 80%), the MSK-IMPACT panel (A), the Illumina TSO500 panel (B), the OncoPrint TML panel (C), and the QIAseq TMB panel (D). **Figure S6.** The accuracy of F1CDx-based TMB estimation varied among cancer types and correlated positively with their TMB levels. **Figure S7.** The mean accuracy of panels comprising 150 to 1000 genes gradually increased but few reached a plateau. (PDF 156 kb)

Abbreviations

ACC: Adrenocortical carcinoma; BLCA: Bladder urothelial carcinoma; BRCA: Breast invasive carcinoma; CESC: Cervical squamous-cell carcinoma and endocervical adenocarcinoma; CHOL: Cholangiocarcinoma; COAD: Colon adenocarcinoma; DLBC: Lymphoid neoplasm diffuse large b-cell lymphoma; ESCA: Esophageal carcinoma; F1CDx: FoundationOne CDx; GBM: Glioblastoma multiforme; HNSC: Head and neck squamous cell carcinoma; ICI: Immune checkpoint inhibitors; KICH: Kidney chromophobe; KIRC: Kidney renal clear cell carcinoma; KIRP: Kidney renal papillary cell carcinoma; LAML: Acute myeloid leukemia; LGG: Brain lower grade glioma; LIHC: Liver hepatocellular carcinoma; LUAD: Lung adenocarcinoma; LUSC: Lung squamous cell carcinoma; Mb: Megabase; MESO: Mesothelioma; NGS: Next-generation sequencing; OV: Ovarian serous cystadenocarcinoma; PAAD: Pancreatic adenocarcinoma; PCPG: Pheochromocytoma and paraganglioma; PRAD: Prostate adenocarcinoma; READ: Rectum adenocarcinoma; SARC: Sarcoma; SKCM: Skin cutaneous melanoma; STAD: Stomach adenocarcinoma; TCGA: The Cancer Genome Atlas; TGCT: Testicular germ cell tumors; THCA: Thyroid carcinoma; THYM: Thymoma; TMB: Tumor mutational burden; UCEC: Uterine corpus endometrial carcinoma; UCS: Uterine carcinosarcoma; UVM: Uveal melanoma; WES: Whole exome sequencing

Acknowledgements

Not applicable.

Authors' contributions

Study concept and design: RX, HW and ZW. Acquisition, analysis, or interpretation of data: All authors. Drafting of the manuscript: All authors. Study supervision: RX and FW. All authors read and approved the final manuscript.

Funding

This work was supported by grants from the National Key R&D Program of China (2018YFC1313300); Natural Science Foundation of Guangdong Province (2017A030313485, 2014A030312015); and Science and Technology Program of Guangdong (2019B020227002).

Availability of data and materials

The TCGA MC3 Public MAF is available at <https://gdc.cancer.gov/about-data/publications/mc3-2017>.

Ethics approval and consent to participate

Ethical approval was waived because we used only publicly available data and materials in this study.

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Author details

¹Department of Medical Oncology, State Key Laboratory of Oncology in South China, Collaborative Innovation Center for Cancer Medicine, Sun Yat-Sen University Cancer Center, Guangzhou 510060, People's Republic of China. ²Bioinformatics Platform, Department of Experimental Research, State Key Laboratory of Oncology in South China, Collaborative Innovation Center for Cancer Medicine, Sun Yat-Sen University Cancer Center, Guangzhou 510060, People's Republic of China.

Received: 11 April 2019 Accepted: 16 July 2019

Published online: 06 August 2019

References

- Rizvi NA, Hellmann MD, Snyder A, Kvistborg P, Makarov V, Havel JJ, et al. Cancer immunology. Mutational landscape determines sensitivity to PD-1 blockade in non-small cell lung cancer. *Science*. 2015;348(6230):124–8.
- Van Allen EM, Miao D, Schilling B, Shukla SA, Blank C, Zimmer L, et al. Genomic correlates of response to CTLA-4 blockade in metastatic melanoma. *Science*. 2015;350(6257):207–11.
- Wang F, Wei XL, Wang FH, Xu N, Shen L, Dai GH, et al. Safety, efficacy and tumor mutational burden as a biomarker of overall survival benefit in chemo-refractory gastric cancer treated with toripalimab, a PD1 antibody in phase Ib/II clinical trial NCT02915432. *Ann Oncol*. 2019. <https://doi.org/10.1093/annonc/mdz197>.
- Rosenberg JE, Hoffman-Censits J, Powles T, van der Heijden MS, Balar AV, Necchi A, et al. Atezolizumab in patients with locally advanced and metastatic urothelial carcinoma who have progressed following treatment with platinum-based chemotherapy: a single-arm, multicentre, phase 2 trial. *Lancet*. 2016;387(10031):1909–20.
- Samstein RM, Lee CH, Shoushtari AN, Hellmann MD, Shen R, Janjigian YY, et al. Tumor mutational load predicts survival after immunotherapy across multiple cancer types. *Nat Genet*. 2019;51(2):202–6.
- Hellmann MD, Ciuleanu TE, Pluzanski A, Lee JS, Otterson GA, Audigier-Valette C, et al. Nivolumab plus Ipilimumab in lung Cancer with a high tumor mutational burden. *N Engl J Med*. 2018;378(22):2093–104.
- Velcheti V, Kim ES, Mekhail T, Dakhlil C, Stella PJ, Shen X, et al. Prospective clinical evaluation of blood-based tumor mutational burden (bTMB) as a predictive biomarker for atezolizumab (atezo) in 1L non-small cell lung cancer (NSCLC): interim B-FIRST results. *J Clin Oncol* 2018;36(15_suppl): 12001–01.
- Allgauer M, Budczies J, Christopoulos P, Endris V, Lier A, Rempel E, et al. Implementing tumor mutational burden (TMB) analysis in routine diagnostics-a primer for molecular pathologists and clinicians. *Transl Lung Cancer Res*. 2018;7(6):703–15.
- Zehir A, Benayed R, Shah RH, Syed A, Middha S, Kim HR, et al. Mutational landscape of metastatic cancer revealed from prospective clinical sequencing of 10,000 patients. *Nat Med*. 2017;23(6):703–13.
- Chalmers ZR, Connelly CF, Fabrizio D, Gay L, Ali SM, Ennis R, et al. Analysis of 100,000 human cancer genomes reveals the landscape of tumor mutational burden. *Genome Med*. 2017;9(1):34.
- Wang Z, Duan J, Cai S, Han M, Dong H, Zhao J, et al. Assessment of blood tumor mutational burden as a potential biomarker for immunotherapy in patients with non-small cell lung Cancer with use of a next-generation sequencing Cancer gene panel. *JAMA Oncol*. 2019. <https://doi.org/10.1001/jamaoncol.2018.7098>.
- Boddy R, Smith G. Statistical methods in practice: for scientists and technologists. Chichester: Wiley; 2009. p. 95–6.
- Ellrott K, Bailey MH, Saksena G, Covington KR, Kandath C, Stewart C, et al. Scalable Open Science approach for mutation calling of tumor Exomes using multiple genomic pipelines. *Cell Syst*. 2018;6(3):271–81.e7.
- Buchhalter I, Rempel E, Endris V, Allgauer M, Neumann O, Volckmar AL, et al. Size matters: Dissecting key parameters for panel-based tumor mutational burden analysis. *Int J Cancer*. 2019;144(4):848–58.
- Endris V, Buchhalter I, Allgauer M, Rempel E, Lier A, Volckmar AL, et al. Measurement of tumor mutational burden (TMB) in routine molecular diagnostics: in silico and real-life analysis of three larger gene panels. *Int J Cancer*. 2019;144(9):2303–12.
- Addeo A, Banna GL, Weiss GJ. Tumor mutation burden-from hopes to doubts. *JAMA Oncol*. 2019. <https://doi.org/10.1001/jamaoncol.2019.0626>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions